

<< Develosil Column >>

TEST REPORT


Packings	Develosil	C8-UG, 5 μm	Batch No.	1Y310719
Column size	Inner diameter	4.6 mm	Length	150 mm
End fitting type	NW			
Mfg. No.	05062077C-RK			

Operating conditions

Mobile phase	:	Acetonitrile:water=70:30
Flow rate	:	1.0 mL/min
Column temp.	:	30 °C
Pressure	:	4.0 MPa
Detection	:	UV 254 nm

Sample (Order of elution)

1	Uracil
2	Naphthalene



Theoretical plate	(Last peak)	$5.54 \times (t_R/W_{0.5})^2 = 12994$
Asymmetry factor	(10% Height of Last peak)	= 1.04
Mobile phase at shipment:	Acetonitrile:water=70:30	

Certificate of Analysis




Develosil[®] C8-UG, 5 μ m Batch# 1Y310719

Analytical Results for Develosil[®] C8-UG, 5 μ m

Analysis of Unbonded Silica Gel		Result
Median Particle Size	[μ m]	5.07
Surface Area	[m ² /g]	296
Pore Volume	[mL/g]	1.16
Average Pore Diameter	[nm]	12.3

Analysis of Bonded Silica Gel		
Total Carbon Content	[%]	10.7

Chromatographic Results for Develosil[®] C8-UG, 5 μ m



Sample:

1. Uracil
2. Caffeine
3. Phenol
4. Amitriptyline
5. Naphthalene

Analytical Conditions;

Column: Develosil C8-UG, 5 μ m (4.6x150mm)
Mobile phase: Methanol/25mM Ammonium Phosphate, pH3.0=70/30
Flow rate: 1.0mL/min
Temperature: 40 $^{\circ}$ C
Detection: UV254nm
Sample: 1.Uracil 2.Caffeine 3.Phenol 4.Amitriptyline 5.Naphthalene
Injection volume: 1.0 μ L

<i>k</i> Naphthalene	Result
Relative retention	0.00
<i>k</i> (Caffeine/Naphthalene)	0.05
<i>k</i> (Phenol/Naphthalene)	0.17
<i>k</i> (Amitriptyline/naphthalene)	0.35
<i>k</i> (Caffeine/Phenol)	0.32
Tailing Factor	
Amitriptyline	1.21

Approved Ikuo Yamamoto

Date: 2019.10.16